

SREE MOOKAMBIKA INSTITUTE OF DENTAL SCIENCES



PADANILAM WELFARE TRUST, V.P.M. HOSPITAL COMPLEX,
PADANILAM, KULASEKHARAM, K.K.DIST. TAMIL NADU, PIN-629161

(Approved by the Govt.of Tamilnadu Recognised by Dental Council of India
and Affiliated to The Tamilnadu Dr.M.G.R. Medical University, Chennai)

Website: smids.sreemookambikainstitute.com

Email id: smidsbds1999@gmail.com

MAINTENANCE LOG BOOK

DEPARTMENT OF PUBLIC HEALTH DENTISTRY

2016

SNO	DATE	Name of item taken for servicing	Quantity	Complaint	Received date	Time of P.T. - 1000	Signature
1	15/1/16	Chair No. 8	1	Light not working	15/1/16		[Signature]
2	3/2/16	Chair No. 5	1	Access pedal not working	5/2/16		[Signature]
3	15/2/16	Micromotor hand piece couriered	1	Not working	15/2/16		[Signature]
4	26/7/16	Fan	1	not working	29/7/16		[Signature]
5	26/7/16	Mouth mirror No. 5	1	Rusted	27/7/16		[Signature]
6	27/7/16	Tube light (near washroom)	1	Not working	29/7/16		[Signature]
7	27/7/16	Autoclave	1	Not working	29/7/16		[Signature]
8	28/7/16	Leakage of Water from tap in the Washroom	1	Leakage	30/7/16		[Signature]

[Handwritten Signature]

Dr. Elizabeth koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam, Kulasekharam-629 101
 K.K. Dist., Tamilnadu



DEPARTMENT OF PUBLIC HEALTH DENTISTRY

2020

7

Date	No of items taken for servicing	Qty	Complaint	Return date	Sign	Staff incharge
2/2/20	Printer of the computer not working	1	Not Working	4/2/20	[Signature]	[Signature]
5/3/20	Aerator not working in Dental Chair No. 4.	1	Not Working	5/3/20	[Signature]	[Signature]
24/11/20	Switch Board near Autoclave not working	1	Not Working	25/11/20	[Signature]	[Signature]
15/12/20	Telephone complaint	1	Not Working	11/12/20	[Signature]	[Signature]

Dr. Elizabeth Koshi MDS
Principal

Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam Kulasekharam-629 181
K.K. Dist., Tamilnadu



DEPARTMENT OF PUBLIC HEALTH DENTISTRY

6 2019

Sr No	Date	No of item taken for Service	Qty	Complained	Repaired date	Sign technician	Sign
21	1/2/19	Computer	1	CPU not working	4/2/19	R. S.	[Signature]
22	3/4/19	Chair No. 7 Water leakage from Hand piece Slot	1	Water leakage from Hand piece Slot	5/4/19	[Signature]	[Signature]
23	6/7/19	Dental Chair No. 2 Aerotec malfunction	1	Aerotec malfunction	9/7/19	[Signature]	Unani Prasad
24	13/9/19	Washbasin Block	1	Blockage in the Washbasin	15/9/19	R. S.	[Signature]

[Signature]

Dr. Elizabeth Koshi MDS
Principal
Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilem Kulasaktharam-628 181
K.K Dist., Tamilnadu



DEPARTMENT OF PUBLIC HEALTH DENTISTRY

2018

SNo	Date	Name of item taken for Repairing	Qty	Complaint	Received date	5 Sign of Staff	Staff Name
14	22/1/18	Micro motor handpiece Contrangled	1	Not Working	25/1/18	<i>[Signature]</i>	Umal Prasad
15	25/1/18	Dental Chair No: 5	1	Pedal not Working	25/1/18	<i>[Signature]</i>	Umal Prasad
16	1/6/18	Telephone Not working	1	Not Working	5/6/18	<i>[Signature]</i>	Umal Prasad
17	4/9/18	Mouth mirror rusted	2	Rusted. (Replace)	6/9/18	<i>[Signature]</i>	<i>[Signature]</i>
18	10/10/18	Dental Chair No: 1	1	Section not Working	12/10/18	<i>[Signature]</i>	<i>[Signature]</i>
19	11/10/18	Bathroom Leakage	1	Water tap leakage	12/10/18	<i>[Signature]</i>	<i>[Signature]</i>
20	20/10/18	Autoclave	1	Repair Not Working	22/10/18	<i>[Signature]</i>	Umal Prasad

[Handwritten Signature]

Dr. Elizabeth koshi MDS
Principal
Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam Kulasekharam-629 181
K.K Dist., Tamilnadu



DEPARTMENT OF PUBLIC HEALTH DENTISTRY

2017

SNO	Date	Name of the item taken for servicing	Qty	Complaint	Returned date	Sign Technician	Staff Sign
9	9/1/17	Dental Chair No. 7	1	Not working light	12/1/17	<i>[Signature]</i>	<i>[Signature]</i>
10	21/1/17	Hot water boiler	1	Short circuit	27/1/17	<i>[Signature]</i>	<i>[Signature]</i>
11	25/2/17	Chair No. 2 Water leakage in Spitter	1	leakage of water in Spitter	28/2/17	<i>[Signature]</i>	<i>[Signature]</i>
12	21/4/17	Dental Chair No. 4	1	Light not functional	19/4/17	<i>[Signature]</i>	<i>[Signature]</i>
13	21/8/17	Tube light near entrance	1	Not Working	26/8/17	<i>[Signature]</i>	<i>[Signature]</i>

[Handwritten Signature]

Dr. Elizabeth koshi MDS
Principal
Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam Kulasekharam-629 18
K.K. Dist., Tamilnadu




DEPARTMENT OF PEDODONTICS AND PREVENTIVE DENTISTRY

2020						2021							
Sl. No.	Date	Name of item taken for servicing	Qty	complaint	returned date	Technician	Sl. No.	Date	Name of item taken for servicing	Qty	complaint	returned date	Technician
29	06-01-2020	Steam cleaners	1	taking long time for boiling	10-01-2020	Ms. Sangeetha	34	21-11-2020	Chair No: 11 (3403/14) Confident	1	light not working Section tube is not functioning	28-11-2020	Ms. Sangeetha
30	21-1-2020	Ultra angled air rotor handpiece (NSK)	1	bus not rotating	25-01-2020	Ms. Sangeetha	35	08-2-2021	Welder with soldering attached (Confident)	1	Obtaining improper welder joint	12-2-2021	Ms. Sangeetha
31	05-02-2020	Chair No: 8 (3403/13) Confident	1	Spillout blocked Airtorator pedal not working	09-02-2020	Ms. Sangeetha	36	15-3-2021	Dental lathe	1	Excessive noise while working	19-3-2021	Ms. Sangeetha
32	20-11-2020	Tubelight in reception area	1	Not functioning	06-11-2020	Ms. Sangeetha	37	19-7-2021	Pulp tester Digital	1	Not functioning Shaving unreliable values	24-7-2021	Ms. Sangeetha
33	14-11-2020	Chair No: 7 (3403/15) Confident	1	Airtorator pedal not working Water not coming from 3 way syringe	15-11-2020	Ms. Sangeetha							

[Signature]

Dr. Elizabeth koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-629 102
 K.K Dist., Tamilnadu



DEPARTMENT OF PEDODONTICS AND PREVENTIVE DENTISTRY

Sl No	Date	Name of item taken for Servicing	Qty	complaint	repaired date	Signature	Sl No	Date	Name of item taken for Servicing	Qty	complaint	repaired date	Signature
19	2/10/2018	Automatic developer	1	light not working obscuring	6/10/2018	Dr. Anand	20	13/05/2019	Endomotor (NSK)	1	handpiece not working	11/05/2019	Dr. Anand
20	10/6/2018	Chair No. 10 (1202/25) confidential	1	3 way syringe not working X-ray View not working	13/07/2018	Dr. Anand	25	20/05/2019	Autoclave (847/92986)	1	noise while cooking	24/05/2019	Dr. Anand
21	10/07/2019	Chair No. 5 (1204/20) confidential	1	h Section not working cable leaking from handpiece slot	13/07/2019	Dr. Anand	26	3/06/2019	Fone (Examination area) (lab area)	1	not functioning	7/06/2019	Dr. Anand
22	21/01/2019	NSK binator hand piece T120021	1	handpiece not functioning	25/01/2019	Dr. Anand	27	17/06/2019	Chair no. 8 (1204/9) confidential	1	cap + don't move movements not possible light not working water leaking	20/06/2019	Dr. Anand
23	1/02/2019	Vibrator (confidential)	1	Excessive sound while working	6/02/2019	Dr. Anand	28	8/01/2019	Amalgamator	1	May and heavy intrusion not done properly	12/01/2019	Dr. Anand


2019




Dr. Elizabeth Koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-629 161
 K.K. Dist., Tamilnadu

DEPARTMENT OF PEDODONTICS AND PREVENTIVE DENTISTRY

2017					2018								
Sl. No	Date	Name of item taken for servicing	Qty	complaint	relieved date	Name and Sign of technician	Sl. No	Date	Name of item taken for servicing	Qty	complaint	relieved date	Name and Sign of technician
8	9/01/2017	Tube light (1157 room)	1	not functioning	12/01/2017	Dr. Sangeetha	12	22/01/2018	chair No. 1 (1204/10) confident	1	up & down movement not possible	25/01/2018	Dr. Sangeetha
9	22/01/2017	Chair No. 2 (1204/9) confident	1	water leakage in spittoon up dental chair light not working	27/01/2017	Dr. Sangeetha	14	22/01/2018	chair no. 9 (3404/9) confident	1	water not coming from spray	25/01/2018	Dr. Sangeetha
10	8/01/2017	Model turnover	1	not properly working	13/02/2017	Dr. Sangeetha	15	22/01/2018	chair no. 11 (3403/11) confident	1	light cover & splash not working rotary handpiece not working		Dr. Sangeetha
11	11/04/2017	Pulp tester digital	1	values obtained are not appropriate	21/04/2017	Dr. Sangeetha	16	1/06/2018	gastroclaw (BMT/42991)	1	not functioning	5/06/2018	Dr. Sangeetha
21	10/01/2017	Apex locator	1	not functioning	26/01/2017	Dr. Sangeetha	17	15/06/2018	light (staff room)	1	not functioning	14/06/2018	Dr. Sangeetha
							18	25/06/2018	light in x-ray room not used	1	not working properly	29/06/2018	Dr. Sangeetha


Dr. Elizabeth Koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam, Kulasekharam-625 151
 K.K. Dist., Tamilnadu



DEPARTMENT OF PEDODONTICS AND PREVENTIVE DENTISTRY

No	Date	Name of item taken for saving	Quantity	Complaint	Returned date	Name of staff in charge	Staff in charge
1	18/1/2016	Amalgamator	1	not working properly	27/1/2016	Santhi Mr. Sangli	Dr. CHINTU
2	18/1/2016	Dental lathe	1	not working properly	2/02/2016	Santhi Mr. Sangli	Dr. Vinayak
3	18/1/2016	contra angled cavitator hand piece (NSK)	1	unable to remove bit	4/02/2016	Santhi Mr. Sangli	Dr. CHINTU
4	3/2/2016	Chair NO: 5 (1204/20) Chair NO: 8 (3403/13) Confident	1	light not working	5/2/2016	Santhi Mr. Sangli	Dr. CHINTU
5	3/2/2016	Chair NO: 6 (1205/17) Chair NO: 12 Confident	1	Microton pedal not working	5/2/2016	Santhi Mr. Sangli	Dr. Vinayak
6	13/06/2016	Microton hand piece contra angled (NSK)	1	not working	17/06/2016	Santhi Mr. Sangli	Dr. CHINTU
7	26/07/2016	Janon Kids waiting area)	1	not working	29/07/2016	Santhi Mr. Sangli	Dr. CHINTU



Dr. Elizabeth Koshi MDS
Principal
Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam, Kulasekharam-629 160
K.K. Dist., Tamilnadu



DEPARTMENT OF ORTHODONTICS

2020

Sl. No.	DATE	NAME OF ITEM TAKEN FOR SERVICES	QUANTITY	NATURE OF COMPLAINT	AMOUNT DATA OF RETURN	NAME & SIGN OF TECHNICIAN	NAME OF (YES) STAFF	RECEIVED DATE	SIGNATURE	DATE
1.	07/02/20	Dental chair no: 5	1	Light sensor not working	11/01/20	Sanjit	Dr. M. Kumar	11/01/20	[Signature]	11/01/20
2.	08/02/20	Maxillary splint	1	not working	01/02/20	Sanjit	Dr. M. Kumar	01/02/20	[Signature]	01/02/20
3.	13/02/20	Aligner	1	Cap damaged	17/02/20	Sanjit	Dr. M. Kumar	11/02/20	[Signature]	11/02/20
4.	01/03/20	Dental chair 4	1	pedal not working	1/03/20	Sanjit	Dr. M. Kumar	1/03/20	[Signature]	1/03/20
5.	27/02/20	SLR digital camera	1	Battery problem	17/02/20	Sanjit	Dr. M. Kumar	17/02/20	[Signature]	17/02/20
6.	03/04/20	Needle destroyer	1	Damaged	03/04/20	Sanjit	Dr. M. Kumar	03/04/20	[Signature]	03/04/20
7.	5/04/20	Dental chair no. 3	1	Broken	09/04/20	Sanjit	Dr. M. Kumar	10/04/20	[Signature]	10/04/20
8.	27/04/20	Aligner no. 3	1	light sensor	2/05/20	Sanjit	Dr. M. Kumar	02/05/20	[Signature]	02/05/20

Dr. Elizabeth koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-629 161
 K.K Dist, Tamilnadu



DEPARTMENT OF ORTHODONTICS

2019

Sl. No.	DATE	NAME OF ITEM TAKEN FOR SERVICING	QUANTITY	NATURE OF DEFECT	DATE IN SERV.	NAME OF STAFF OR TECHNICIAN	NAME OF CLINIC	RESERVED BY	Signature	Date
1.	30.07.19	Dental chair Unit No. 5	1	Pedal not working	2/8/19	Sanjiv	Dr. Anil's Ni.	2/8/19	[Signature]	2/8/19
2.	7/8/19	Set of orthodontic instruments	6	needle leakage	11/8/19	Sanjiv	Dr. Anil Kumar	11/8/19	[Signature]	11/8/19
3.	23.09.19	Refrigerator	1	water leakage	25.09/19	Sanjiv	Dr. Anil Kumar	25.09/19	[Signature]	25/9/19
4.	1/9/19	Free place articulator	1	fully damaged	5/9/19	Sanjiv	Dr. Anil Kumar	5/9/19	[Signature]	5/9/19
5.	20-10-19	Dental chair unit no - 2	1	Light Not working	2/10/19	Sanjiv	Dr. Anil Kumar	2/10/19	[Signature]	2/10/19
6.	12.11.19	LCD Projector	1	Remote control not functioning	16.11.19	Sanjiv	Dr. Anil Kumar	16.11.19	[Signature]	16/11/19
7.	12/12/19	Polishing Lath	1	Not working	16/12/19	Sanjiv	Dr. Anil Kumar	17/12/19	[Signature]	17/12/19
8.	16/12/19	Dental Chair 1	1	light sensor not working	22/12/19	Sanjiv	Dr. Anil Kumar	22/12/19	[Signature]	22/12/19

D

Dr. Elizabeth koshi MDS
Principal
Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanitam Kulasekharam-629 16
K.K. Dist., Tamilnadu



DEPARTMENT OF ORTHODONTICS

2018									
S.No	DATE	Name Of Asset Taken For SERVING	Quantity	Nature Of Complaint	APPROVAL DATE OF RETN	Name Of Staff	Name Of (In)	Staff No	Staff in charge of Dept
						TECHNICIAN	STAFF	REMOVED IN	
1	05.01.18	Dental Chair Unit no. 3	1	Partial not activating	05.01.18	Srjiti	Dr. Sait Kumar	07.01.18	
2	16.2.18	Tracing Table	1	Fully Damaged	19.02.18	Srjiti	Dr. David Thomas Sunny	19.02.18	
3	22/3/18	Hand bitamer	1	Not working	27/3/18	Srjiti	Dr. Sait Kumar	27/3/18	
4	17/4/18	Dental chair Unit no. 4	1	Failure not working	20/4/18	Srjiti	Dr. Sait Kumar	21/4/18	
5	11.06.18	Computer software for cephalometric analysis	1	System hanging	14.06.18	Srjiti	Dr. Anil S Nair	16.06.18	
6	11/06/18	Polishing lathe	1	Motor Not working	15/8/18	Srjiti	Dr. Sait Kumar	15/8/18	
7	16/10/18	Dental chair-7	1	Light not working	25/10/18	Srjiti	Dr. Anil S Nair	25/10/18	
8	16.11.18	DR Digital Camera	1	Lens Broken	21.11.18	Srjiti	Dr. Sait Kumar	23/11/18	

Dr. Elizabeth Koshi MDS
Principal
Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam, Kulasekharam-629 106
K.K. Dist., Tamilnadu



DEPARTMENT OF ORTHODONTICS

2017									
No.	DATE	NAME OF ITEM TAKEN FOR SERVICING	QUANTITY	NATURE OF COMPLAINT	DATE OF ITEM	NAME OF CLINIC	NAME OF HEAD	DATE OF VISIT	SIGNATURE
1.	16/2/17	Dental chair unit No. 1	1	Chair not working	16/2/17	Sanjit	Dr. Anil Kumar	16/2/17	[Signature]
2.	23/2/17	Dental chair unit no. 4	1	Acrylic problem	27/2/17	Sanjit	Dr. Anil Kumar	27/2/17	[Signature]
3.	15/4/17	X-ray viewer	1	Not working	17/4/17	Sanjit	Dr. Anil Kumar	17/4/17	[Signature]
4.	02/05/17	Dental chair unit no. 7	1	Chair not working	06/05/17	Sanjit	Dr. Anil Kumar	06/05/17	[Signature]
5.	18/1/17	Lab Micrometer	1	Damage of dial	5/2/17	Sanjit	Dr. Anil Kumar	5/2/17	[Signature]
6.	26/1/17	Self orthodontic plier	4	Blunt tip	2/2/17	Sanjit	Dr. Anil Kumar	2/2/17	[Signature]
7.	24/1/17	Hand held scaler	1	Scalor not working	23/1/17	Sanjit	Dr. Anil Kumar	23/1/17	[Signature]
8.	9/12/17	Dry heat sterilizer	1	Not working	12/12/17	Sanjit	Dr. Anil Kumar	12/12/17	[Signature]
9.	13/12/17	Dental chair unit No. 6	1	Light sensor problem	16/12/17	Sanjit	Dr. Anil Kumar	16/12/17	[Signature]

Dr. Elizabeth Koshi MDS
Principal
Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam Kulasekharan-629 101
K.K. Dist., Tamilnadu



DEPARTMENT OF ORTHODONTICS

21					2016					
S.N.	DATE	NAME OF ITEM TAKEN FOR SERVICING	QUANTITY	NATURE OF COMPLAINT	DATE OF REPAIR	NAME & DESIG. OF TECHNICIAN	NAME OF (DOC) STAFF	DATE REPAIRED	DR. NAME	Signature
1.	08.01.16	Dental chair unit No. 2	1	Swivel Not working	21.01.16	Sajith	Dr. Anil Kumar	21.01.16	Dr. Anil Kumar	[Signature]
2.	08.12.16	Autoclave for bulk instrument Sterilization Vacuum	1	Damaged Coil	12.02.16	Sajith	Dr. Anil Kumar	12.02.16	Dr. Anil Kumar	[Signature]
3.	25.03.16	Dental chair unit No. 1	1	Roller Not working	29.02.16	Sajith	Dr. Anil Kumar	29.02.16	Dr. Anil Kumar	[Signature]
4.	14.4.16	Ultrasonic scaler	1	Not working	16.4.16	Sajith	Dr. Anil Kumar	16.4.16	Dr. Anil Kumar	[Signature]
5.	20.5.16	Dental chair unit No. 3	1	Light not working	23.5.16	Sajith	Dr. Anil Kumar	23.5.16	Dr. Anil Kumar	[Signature]
6.	15.8.16	Welding Machine	1	Not working	17.08.16	Sajith	Dr. Anil Kumar	17.08.16	Dr. Anil Kumar	[Signature]
7.	12.09.16	Light curing unit	2	Battery problem	14.09.16	Sajith	Dr. Anil Kumar	14.09.16	Dr. Anil Kumar	[Signature]
8.	21.10.16	Dental chair unit No. 5	1	Swivel Not working	31.10.16	Sajith	Dr. Anil Kumar	31.10.16	Dr. Anil Kumar	[Signature]


Dr. Elizabeth Koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam, Kulasekharam-629 101
 K.K Dist., Tamilnadu



DEPARTMENT OF PROSTHODONTICS

Date	Name of Item Intended for repairs	Quantity	Nature of Complaint	Appropriate Date of return	Name and Signature of technician	Name and Signature of MOB/Staff	Date of Return/ Repair	Signature MOB
31/6/2020	Dental chair (20.3)	②	Not working	10/6/2020	Ravi	[Signature]	10/6/2020	
4/6/2020	Fan of hall light	①	Not working	15/6/2020	Vijay	[Signature]	20/6/2020	
12/6/2020	Air conditioner in Implant room	1	Not working	25/6/2020	Johanna	[Signature]	20/6/2020	
31/7/2020	Gas burner	3	Not working	15/7/2020	Mani	[Signature]	26/7/2020	
14/8/2020	Die cutting machine	1	Not working	24/8/2020	Vijay	[Signature]	29/8/2020	
17/9/2020	High speed Lathe	①	Not working	20/9/2020	[Signature]	[Signature]	12/9/2020	
21/11/2020	Fan sitting height	2	Not working	20/11/2020	Vijay	[Signature]	28/11/2020	
01/12/2020	Dental chair 1, 6	2	not moving up & down	10/11/2021	Ravi		15/11/2021	


Dr. Elizabeth Koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-629 101
 K.K. Dist., Tamilnadu



DEPARTMENT OF PROSTHODONTICS

18 19

Date	Name of the Instructor	Quantity	Name of Company	Material Name and Quantity	Material Name and Quantity	Date of Report	Signature of Instructor
			No water coming	15/5/2020	Ravi	15/5/2020	[Signature]
			No sound	15/5/2020	Vijay	10/5/2020	[Signature]
			not working	15/5/2020	Vijay	15/5/2020	[Signature]
			not working	20/5/2020	Ravi	14/5/2020	[Signature]
			Not working	15/5/2020	Vijay	15/5/2020	[Signature]
			not working w/r down	15/5/2020	Ravi	20/5/2020	[Signature]
			Not working	15/5/2020	Vijay	20/5/2020	[Signature]
			not working & sound coming	20/5/2020	Ravi	25/5/2020	[Signature]


Dr. Elizabeth Koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-629
 K.K. Dist., Tamilnadu



DEPARTMENT OF PROSTHODONTICS

1A				1B				
Date	Name of Item taken for repairs	quantity	Order of supply	Appropriate date of order	Name and signature of Instructor	Name and Signature of MOS/Staff	Date of taken/ Suppl	Signature of Recipient
21/1/2019	Dental chair 8, 9	2	light not working	20/1/2019	<i>[Signature]</i>	<i>[Signature]</i>	20/1/2019	<i>[Signature]</i>
20/1/2019	Acrylic base (3)	1	Not working	16/1/2019	<i>[Signature]</i>	<i>[Signature]</i>	28/1/2019	<i>[Signature]</i>
14/1/2019	Tube light	3	Not working	28/1/2019	<i>[Signature]</i>	<i>[Signature]</i>	28/1/2019	<i>[Signature]</i>
22/1/2019	Incenti Stiffness	2	Not working	20/1/2019	<i>[Signature]</i>	<i>[Signature]</i>	28/1/2019	<i>[Signature]</i>
11/1/2019	Dental chair 1, 2, 18	3	Spillins water control not working	20/1/2019	<i>[Signature]</i>	<i>[Signature]</i>	28/1/2019	<i>[Signature]</i>
17/1/2019	Dental chair 13	1	No. inside water supply	20/1/2019	<i>[Signature]</i>	<i>[Signature]</i>	28/1/2019	<i>[Signature]</i>
21/1/2019	Instrument finance		Not working	21/1/2019	<i>[Signature]</i>	<i>[Signature]</i>	10/1/2019	<i>[Signature]</i>
12/1/2019	Dental chair 13, 4, 18	3	Not working	10/1/2019	<i>[Signature]</i>	<i>[Signature]</i>	11/1/2019	<i>[Signature]</i>

[Signature]
Dr. Elizabeth koshi MD
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-628 1
 K.K Dist., Tamilnadu




DEPARTMENT OF PROSTHODONTICS

UG 2018

Date	Name of items	Quantity	Nature of the items	Appropriate Crown and	Material	Signature of	Date of	Signature of
				Internal	Signature of	Departmental	University /	MS
				material	Department	Office/Case	Program	
3/1/2018	Lab 2	1	Accuracy	1/1/2018	P. S.	[Signature]	16/1/2018	[Signature]
10/1/2018	Dental chair 12, 4	2	Speller with control and working	2/1/2018	P. S.	[Signature]	29/1/2018	[Signature]
18/1/2018	Chair 1201	1	Control working	8/1/2018	Harish	[Signature]	20/1/2018	[Signature]
8/1/2018	Dental chair 10, 8	2	Light and working	15/1/2018	V. S.	[Signature]	17/1/2018	[Signature]
10/1/2018	Tubalgit	4	Not working	15/1/2018	V. S.	[Signature]	15/1/2018	[Signature]
10/1/2018	Adolam	1	Not working	20/1/2018	Harish	[Signature]	17/1/2018	[Signature]
16/1/2018	Chair 9, 13, 20	3	Light and working	30/1/2018	P. S.	[Signature]	27/1/2018	[Signature]
11/1/2018	Chair 11, 16	2	No Patient with lamp	20/1/2018	P. S.	[Signature]	22/1/2018	[Signature]




Dr. Elizabeth Koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-629 101
 K.K. Dist., Tamilnadu

DEPARTMENT OF PROSTHODONTICS

Oct 2017

Date	Name of item Reason for reporting	Quantity	Nature of Complaint	Appropriate date of return	Name and Signature of technician	Name and Signature of non-tech	Date of Return/Repair	Signature of non-tech
17/8/2017	Dental chair 8, 10	2	Not working	20/8/2017	[Signature]	[Signature]	22/8/2017	[Signature]
22/8/2017	Dental Lake 1, 2	2	Not working	25/8/2017	[Signature]	[Signature]	27/8/2017	[Signature]
23/9/2017	Auto chair	1	Not working	11/10/2017	[Signature]	[Signature]	10/10/2017	[Signature]
11/10/2017	Scalizer	1	Not working	20/10/2017	[Signature]	[Signature]	22/10/2017	
4/6/2017	Tubelights	4	Not working	15/6/2017	[Signature]	[Signature]	16/6/2017	
8/8/2017	High speed Lab	1	Not working	22/8/2017	[Signature]	[Signature]	19/8/2017	
21/12/2017	Dental Chair 21	1	Light not working	10/1/2018	[Signature]	[Signature]	12/1/2018	[Signature]
20/11/2017	Parrot	1	Not working	30/11/2017	[Signature]	[Signature]	5/12/2017	
11/12/2017	Fan	2	Not working	20/11/2017	[Signature]	[Signature]	28/12/2017	[Signature]

[Handwritten signature]



Dr. Elizabeth Koshi MDS
Principal
Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam Kulasekheram-629 181
K.K. Dist., Tamilnadu

DEPARTMENT OF PROSTHODONTICS

10/1 2017

Date	Name of Item (Refer to reference)	Quantity	Nature of Complaint	Appraised date of repair	Name and Signature of Technician	Name and Signature of MOB/Unit	Date of Repair/Work	Signature of MOB
14/1/2017	DC ending Machine	1	Not working	20/1/2017	P.S.	[Signature]	5/6/2017	T. S. S.
2/2/2017	Dental chair 4, 2	111	Function not working [unclear]	20/3/2017	P.S.	[Signature]	16/3/2017	
22/2/2017	Machine	1	Not working	5/5/2017	[Signature]	[Signature]	20/4/2017	T. S. S.
9/5/2017	Dental chair 11	1	Low number pressure	15/5/2017	P.S.	[Signature]	20/6/2017	T. S. S.
17/6/2017	Dental chair 5	1	No rotation water supply	30/6/2017	P.S.	[Signature]	24/11/2017	T. S. S.
2/6/2017	Sub-light	2	Not working	20/6/2017	[Signature]	[Signature]	5/1/2018	



Dr. Elizabeth Koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam, Kulasekharam-629 101
 K.K. Dist., Tamilnadu

DEPARTMENT OF PROSTHODONTICS

2		UG 2016		3			
Date	Name of the item taken for repairing	Quantity	Nature of complaint	Associated item and date of signature of technician	Person and signature of who/WH	Date of repair/repair	Signature of who/staff
18/3/2016	Dental chair 2, 10, 18		Light not working	24/2/2016	P.S.	27/2/2016	up to 1 hr
3/3/2016	raygun (A) No. 1 in table (A)	1	Not working	24/2/2016	P.S.	28/3/2016	
30/3/2016	Dental chair 8, 23		Water not coming	15/4/2016	P.S.	12/4/2016	up to 6 hrs
11/5/2016	High speed lab (C)	1	Not working	20/11/2016	Uma	9/5/2016	
15/7/2016	Current furnace		Not heating	15/2/2016	P.S.	6/8/2016	
19/8/2016	2 tube light fixture	2	Not working (the software class)	24/8/2016	Uma	24/8/2016	
7/10/2016	Dental chair 4, 16, 27	3	No water coming Water light not working	15/10/2016	P.S.	15/10/2016	up to 6 hrs
11/12/2016	Autoclave 1	1	Not working	22/12/2016	Uma	22/12/2016	
18/12/2016	Dental chair 14, 23	2	No water in the chair	31/12/2016	P.S.	10/1/2017	



Dr. Elizabeth Koshi MDS
Principal
Sree Monkambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam Kulasekharam-629 101
K.K. Dist., Tamilnadu

DEPARTMENT OF PROSTHODONTICS

Pg 2016

Date	Name of item taken for repairing	Quantity	Nature of Complaint	Approximate date of repair	Name and Signature of technician	Name and Signature of MOD/STAFF	Date of Repair/issue	Signature of MOD/STAFF
22/3/2016	Dental chair 5,7	2	No water in water.	20/3/2016	<i>Pr</i>	<i>Pr</i>	23/3/2016	
28/4/2016	Aluminum in implant room	1	not working	01/5/2016	<i>Pr</i>	<i>Pr</i>	15/5/2016	f.l.h
12/5/2016	Dental chair 10	1	Not moving up/down	20/5/2016	<i>Pr</i>	<i>Pr</i>	15/5/2016	
9/6/2016	Autoclave	1	not working	15/6/2016	<i>Pr</i>	<i>Pr</i>	16/6/2016	
30/6/2016	Burner	1	No flame in the burner	14/7/2016	<i>Pr</i>	<i>Pr</i>	20/7/2016	f.l.h
23/7/2016	3.900 lights	3	not working	20/7/2016	<i>Pr</i>	<i>Pr</i>	24/7/2016	
15/9/2016	Dental chair 8	1	Section not working	20/9/2016	<i>Pr</i>	<i>Pr</i>	25/9/2016	f.l.h
11/12/2016	Dental chair 14	2	Light not working	20/12/2016	<i>Pr</i>	<i>Pr</i>	24/12/2016	




Dr. Elizabeth koshi MDS
Principal
Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam, Kulasekharam-629 161
K.K. Dist., Tamilnadu

DEPARTMENT OF ORAL PATHOLOGY

2020

Sl. No.	Date	Name of class / Lab / Seminar	Qty	Complaints		Reported by	Signature	Date
				UOI	PCN			
1	8/1/2020	Switch board Dept library	1		Not working	26/1/2020	[Signature]	8
2	11/2/2020	Lubricant Lx200 microscope Sn No-19101 1920	1		objective (lens) not clear	24/2/2020	[Signature]	8
Department closed from 21 st Mar 2020 to 29 th Sep 2020								
3	6/10/2020	Fan (Pb room)	1		Fan not working	22/10/2020	[Signature]	8
4	18/11/2020	Tubelight (Lab)	1		light not working	29/12/2020	[Signature]	8
5								




Dr. Elizabeth koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-626 161
 K.K Dist., Tamilnadu

DEPARTMENT OF ORAL PATHOLOGY

S.No	Date	Name of item including Qty.	Complaints		Repaired on	Sign of lab/clin	Sign of staff	S.No	Date	Name of item including Qty.	Complaints		Repaired on	Sign of lab/clin	Sign of staff
			On	By							On	By			
1	2/1/2019	Computer (space/1000 room) CPU-	Computer not working		24/1/2019										
2	5/1/2019	Printer Calcutage	Printer not working		20/1/2019										
3	12/2/2019	Fan (V6 Lab)	Fan not working		29/2/2019										
4	14/6/2019	Labmed Lx200 microscope Sv No:- 141074467	objective (Lx200) not working		28/6/2019										
5	24/7/19	Tube light (Seminar room)	light not working		16/7/19										
6	11/9/19	Printer Calcutage	Printer not working		26/10/2019										





Dr. Elizabeth Koshi
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-629 101
 K.K Dist., Tamilnadu

DEPARTMENT OF ORAL PATHOLOGY

8		2018		7		
S.No	Date	Name of Item	Complaints		Date of Repair	Status
			No	Yes		
7	4/9/2018	Fan (Seminar room-1)	Not working		20/7/2018	✓
10	2/10/18	Labored Lx 100 Sx No-1107 49E-dark field micro scope	objective (10x) not clear		9/11/2018	✓
11	20/11/18	Light (Staff room)	Tube light not working		4/12/2018	✓
1	7/1/2018	Labored Lx 100 micro scope Sx No-1149 600X Phase contrast microscope	condenser ring not working		22/1/2018	✓
2	13/3/2018	Labored Lx 100 camera not see	Camera not working		22/3/2018	✓
3	7/6/2018	Fan (Prof. room room)	Not working		19/6/2018	✓
4	20/7/2018	Labored Lx 100 Sx No 141214432 microscope	Microscope not working		8/7/2018	✓
5	18/9/2018	Tube light (110 Lab)	Light not working		9/10/2018	✓




Dr. Elizabeth koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilem Kulasekharam-629 161
 K.K. Dist., Tamilnadu

DEPARTMENT OF ORAL PATHOLOGY

Sl. No.	Date	Name of item taken for working Qty	Complaint Qty	Repair on	Sign of lab	Sign of staff	2017		Repair on	Sign of Lab/Staff	Sign of Staff
							Ute	PSI			
1	11/1/2017	Iron/Seam (room)	1	Not working	28/1/2017	OK	OK	Not working	28/1/2017	OK	OK
2	21/1/2017	Labeled (Lab)	1	Not working	28/1/2017	OK	OK	Objective (100x) not clear	16/3/17	OK	OK
3	21/1/2017	Labeled microscope Lx 200 S/N No-1410144 66	1	Objective (100x) not clear	28/1/2017	OK	OK	Microscope not working	2/2/17	OK	OK
4	22/2/2017	Unlabeled compound microscope	1	not working	28/1/2017	OK	OK	Fan (Vie room)	20/3/17	OK	OK
5	20/4/2017	Labeled Lx 200 S/N No-1410144	1	Microscope not working	28/1/2017	OK	OK	Microscope not working	29/4/17	OK	OK
6	4/5/2017	Tube light (Vie Lab)	1	Tube light not working	28/1/2017	OK	OK	Tube light not working	28/5/17	OK	OK
7	17/6/2017	Computer (MOD. Room)	1	not working	28/1/2017	OK	OK	Camera not working	10/8/17	OK	OK
8	20/7/17	Labeled (Lx 200) microscope S/N No 9144100	1	Camera not working	28/1/2017	OK	OK	Camera not working	10/8/17	OK	OK




Dr. Elizabeth Koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-628 100,
 K.K Dist., Tamilnadu

DEPARTMENT OF ORAL PATHOLOGY

3

2016

No	Date	Name of item taken for servicing	Qty	Complaint		Repaired on	Sign of technician on	Sign of staff
				Qty	PLG			
1	18/02/16	Labomed microscope Sr No - 141074411	1		objective (40x) not clear	26/02/16		
2	26/02/16	Adelta microscope Sr No. AV009-09012 TDM 17 Polarizer and analysis box	1		Microscope not working	10/02/16		
3	23/3/16	Dental chair	1		light not working	31/03/16		
4	14/4/16	Microtome	1		Microtome not working	29/4/16		
5	5/5/2016	Labomed Lx500 9144600 (Phase Contrast microscope)	1		objective 10x not clear	19/5/2016		
6	19/5/2016	Labomed Lx200	1		Objective (10x) not clear	28/5/2016		
7	15/6/2016	Labomed Microscope Lx200 Sr No - 141074469 (Condenser ring)	1		Microscope not working	30/6/2016		
8	7/7/2016	Dental chair	1		Spray Syringe not working	21/7/16		




Dr. Elizabeth Koshi MDS
Principal

Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam Kulasekharam-629 189
K.K. Dist., Tamilnadu

Oral Medicine and Radiology

		2019		Complaint		Repair Tech	Staff
Date	Name of item taken for service	Qty	UG	PG	on	Sign	Sign
19/03/19	Stemliner	1		1 (not working)	19/03/19	[Signature]	[Signature]
24/03/19	OPG	1		1 (Control panel display error)	24/03/19	[Signature]	[Signature]
22/04/19	Dental chair 12, 13, 14, 15	4	2	2 (Air and water blockage)	30/04/19	[Signature]	[Signature]
04/05/19	OPG Regular Service	1		1 Regular Service	24/05/19	[Signature]	[Signature]
15/05/19	AC	1		1 (low cooling - freezing)	19/05/19	[Signature]	[Signature]
19/05/19	Remote control (AC)	2		2 (Battery change)	19/05/19	[Signature]	[Signature]
21/02/19	Tube light	4		4	21/02/19	[Signature]	[Signature]
14/02/19	UPS	2		2 (UPS not working)	16/02/19	[Signature]	[Signature]
16/02/19	Printer (6-page)	2		2 (Cartridge change)	16/02/19	[Signature]	[Signature]
33/11/19	Autoclave	1		1 (Cal change)	05/12/19	[Signature]	[Signature]
10/12/19	Viewbox	1		1 (light change)	10/12/19	[Signature]	[Signature]
23/12/19	CBCT	1		1	04/01/2020	[Signature]	[Signature]

		2020		Complaint		Repair Tech	Staff
Date	Name of item taken for service	Qty	UG	PG	on	Sign	Sign
05/03/20	Dental chair 12, 13	2	2	2 (water blockage)	05/03/20	[Signature]	[Signature]
05/03/20	sterilizer	2	2	2 (lead coil change)	07/03/20	[Signature]	[Signature]
15/03/20	Tube light	2	2		15/03/20	[Signature]	[Signature]
16/03/20	Electric pulp tester	3		3 (Battery change)	16/03/20	[Signature]	[Signature]
23/04/20	AC	2		2 (Regular Service)	23/04/20	[Signature]	[Signature]
12/05/20	Internet modem	1		1 (not working)	12/05/20	[Signature]	[Signature]
21/05/20	CBCT + OPG	2		2 (Regular Service)	21/05/20	[Signature]	[Signature]
15/06/20	Dental chair 12, 14, 15	3	2	1 (Air and water blockage)	15/06/20	[Signature]	[Signature]
21/06/20	Dental chair 9, 10, 11, 14, 15	5	3	2 (light not working, Air and water blockage)	21/06/20	[Signature]	[Signature]
01/07/20	OPG Printer	1		1 (firmware)	05/07/20	[Signature]	[Signature]
01/07/20	CPU + Printer	2		2	01/07/20	[Signature]	[Signature]



Dr. Elizabeth Jebish MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-628 101,
 K.K Dist., Tamilnadu

Oral Medicine and Radiology

2017		Complaint		Repair on	Rec. Sign	Man. Sign
Date	Name of Item Taken for Servicing Qty	Uls	Pls			
12/07/17	Jopa Machine 1	Board change 1		18/10/17		
21/07/17	AC	Reduced cooling effn 1		23/07/17		
22/07/17	Remote Control AC 1	Not working		22/07/17		
26/07/17	Regulator AC Service 2		2	26/07/17		
04/08/17	Processing tank Jailed tank 2	2 change		04/08/17		
06/08/17	OPG Machine 1	Board component		16/11/17		
12/08/17	Dental chair 3	2	1	12/08/17		
18/08/17	Selectric Pulp tester 2	Battery change 2		18/08/17		
29/08/17	Tubelight 3	3 (light not working)		06/09/17		
01/09/17	BP Apparatus 1	Not working		06/09/17		
02/09/17	OPG Regular Service	Regular Service		18/10/17		
02/09/17	Moulam 1	Service Connection error 3		05/09/17		
02/09/17	K7 Antivirus 1	Antivirus updated		06/09/17		
10/09/17	Safelight 1	Light not working		12/09/17		

2018		Complaint		Repair on	Rec. Sign	Man. Sign
Date	Name of Item Taken for Servicing Qty	Uls	Pls			
12/08/18	Computer Printer 2			18/08/18		
21/08/18	Tube light 4	2	2	23/08/18		
20/08/18	Dental chair 2	1	1	20/08/18		
30/08/18	OPG Printer not working			02/09/18		
02/09/18	Artelase 1	Not working		14/09/18		
13/09/18	OPG Service	Regular Service		13/09/18		
21/09/18	chair - 1,3,5 3	2	1	22/09/18		
26/09/18	cpu 1			28/09/18		
06/10/18	chair, 1,3,5 3	3	3	06/10/18		
13/10/18	Tubelight 3			23/10/18		
02/10/18	OPG Printer 1	Film loading error		02/10/18		
11/10/18	Safelight 1			14/10/18		
02/10/18	Stentrac unit 1	1		02/10/18		



Dr. Elizabeth koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-629 189
 K.K Dist., Tamilnadu

Oral Medicine and Radiology

2016

Sl. No Date	Name of Item Taken for servicing	Qty	COMPLAINT		Repaired on	Tech Sign	MOD/ Sd/Sl Sign
			Un	Pst			
1.							
4/4/16	Fan not working	2	1	1	04/04/16	[Signature]	[Signature]
12/4/16	Chair no. 08		water blockage		12/04/16	[Signature]	[Signature]
22/4/16	Ops	1		Ops machine not working	22/05/16	[Signature]	[Signature]
18/05/16	printer	1		Ops printer not working	23/05/16	[Signature]	[Signature]
20/05/16	Light	2	Dark room light		22/05/16	[Signature]	[Signature]
22/05/16	Safe light	1	Dark room		02/06/16	[Signature]	[Signature]
25/05/16	Chair 13, 17	2	Light not working		25/05/16	[Signature]	[Signature]
2/10/16	Ops Regular Service			Regular service	12/10/16	[Signature]	[Signature]
4/12/16	Light	1		Not working	16/12/16	[Signature]	[Signature]
22/02/17	computer ops	3		Not working	05/02/17	[Signature]	[Signature]
21/02/17	Chair 4	1	Air Blockage		13/02/17	[Signature]	[Signature]
2/03/17	working stool	4	2	2	12/03/17	[Signature]	[Signature]
6/03/17	Viewbox	2	1	light not working	15/03/17	[Signature]	[Signature]
5/03/17	Regular Service OPU	1			15/03/17	[Signature]	[Signature]
5/03/17	Dental chair	3		Air blockage	20/03/17	[Signature]	[Signature]
10/03/17	view box	1	1	light not working		[Signature]	[Signature]



Dr. Elizabeth Koshi MDS
Principal

Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam Kulasekharan-629 100
K.K. Dist., Tamilnadu

DEPARTMENT OF ORAL & MAXILLOFACIAL SURGERY

S.No	Date	Name of the item taken for servicing	Qty	Complaint	Repaired on	Sign of	Sign of
1	4/1/20	Chair - 3404/5 (Confident)	1	light not working	7/1/20	<i>[Signature]</i>	<i>[Signature]</i>
2	11/3/20	Morshon - 4 Micromorbe Serial no - 9226341	1	Not working	19/3/20	<i>[Signature]</i>	<i>[Signature]</i>
3	8/6/20	Chair - 1204/7 (Confident)	1	water not coming	12/6/20	<i>[Signature]</i>	<i>[Signature]</i>
4	11/8/20	Unique clare C-79B (Confident)	1	not working	15/8/20	<i>[Signature]</i>	<i>[Signature]</i>
5	15/10/20	Chair - 3401/20 (Confident)	1	Foot pedal not working	19/10/20	<i>[Signature]</i>	<i>[Signature]</i>
6	1/12/20	Chair Surgic XT Model - NE111 physio dispenser	1	Foot Control not working	4/12/20	<i>[Signature]</i>	<i>[Signature]</i>



Dr. Elizabeth koshi MOS
 / Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam, Kulasekharam-629 101
 K.K Dist., Tamilnadu

DEPARTMENT OF ORAL & MAXILLOFACIAL SURGERY

S.No	Date	Name of the item taken for servicing	Qty	Complaint	Repaired on	Sign of Technician	Sign of Staff
1	2/1/19	chair - 1204/4 (Confident)		light not working	15/1/19	<i>[Signature]</i>	<i>[Signature]</i>
2	9/3/19	Masathun - 3 Micrometer serial no - 018789		Not working	12/3/19	<i>[Signature]</i>	<i>[Signature]</i>
3	10/6/19	chair - 3405/12 (Confident)		Foot pedal not working	13/6/19	<i>[Signature]</i>	<i>[Signature]</i>
4	17/7/19	chair - 3405/19 (Confident)		light not working	20/7/19	<i>[Signature]</i>	<i>[Signature]</i>
5	1/11/19	NSK Micro Saw Handpiece SGR2-E/SG02-E/SGT2-E		not working	5/11/19	<i>[Signature]</i>	<i>[Signature]</i>
6	20/12/19	chair - 3405/5 (Confident)		Water not coming	23/12/19	<i>[Signature]</i>	<i>[Signature]</i>



Dr. Elizabeth koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-629 101
 K.K. Dist., Tamilnadu

DEPARTMENT OF ORAL & MAXILLOFACIAL SURGERY

2018

S.No	Date	Name of the item taken for servicing	Qty
1	6/1/18	chair - 3401/18 (Confident)	1
2	12/3/18	ultraviolet chamber	1
3	16/5/18	chair - 3401/5 (Confident)	1
4	7/7/18	Unique clare C-79 B (Confident)	1
5	15/11/18	chair - 1204/8 (Confident)	1
6	8/12/18	Marathon Straight handpiece ES-8	1
7	11/12/18	chair - 3401/18 (Confident)	1

Complaint	Repaired on	Sign of Technician	Sign of Staff
Light not working	9/1/18	<i>[Signature]</i>	<i>[Signature]</i>
not working	15/3/18	<i>[Signature]</i>	<i>[Signature]</i>
Foot pedal not working	18/5/18	<i>[Signature]</i>	<i>[Signature]</i>
not working	11/7/18	<i>[Signature]</i>	<i>[Signature]</i>
Light not working	17/11/18	<i>[Signature]</i>	<i>[Signature]</i>
not working	11/12/18	<i>[Signature]</i>	<i>[Signature]</i>
Foot Control not working	10/2/18	<i>[Signature]</i>	<i>[Signature]</i>



Dr. Elizabeth koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam, Kulasekharam-629 181
 K.K. Dist., Tamilnadu

DEPARTMENT OF ORAL & MAXILLOFACIAL SURGERY

6

S.No	Date	Name of the item taken for Servicing	Qty
1	4/2/17	chair - 3401/4 (Confident)	1
2	11/4/17	chair - 3402/3 (Confident)	1
3	10/6/17	chair - 3404/1 (Confident)	1
4	21/8/17	Unique clare C-796 (Confident)	1
5	5/10/17	Dentsuge straight hand pie 18E2627	1
6	14/12/17	Apl PM-Lowal Straight hand pie SN-619156603	1

2017

7

U/G	Complaint	Repaired on	Sign of technician	Sign Staff
	Foot pedal not working	7/2/17	Sanghi	SS
	Light not working	14/4/17	Sanghi	SS
	light not working	13/6/17	Sanghi	SS
	not working	23/8/16	Sanghi	SS
	Not working	9/10/16	Sanghi	SS
	Not working	18/12/17	Sanghi	SS



Dr. Elizabeth Koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-629 101
 K.K Dist., Tamilnadu

DEPARTMENT OF ORAL & MAXILLOFACIAL SURGERY

2016

No	Date	Name of the item taken for servicing	Qty	Complaint	Repair on	Signd technician	Signd Staff
1.	9/1/16	Chair - 3401/12 (Confident)	1	Light is not working	13/1/16	[Signature]	[Signature]
2.	10/3/16	Unique clare C-79B (Confident)	1	Not working	12/3/16	[Signature]	"
3.	12/6/16	Chair - 3405/9 (Confident)	1	Suction not working	16/6/16	[Signature]	[Signature]
4.	16/7/16	Chair - 3401/5 (Confident)	1	Foot Control not working	19/7/16	[Signature]	"
5.	23/9/16	Sanyang Straight Handpiece SDF-ES6	1	Not working	26/9/16	[Signature]	"
6.	19/11/16	Chair - 1204/3 (Confident)	1	Water not coming	22/11/16	[Signature]	[Signature]
7.	10/12/16	Chair - 3405/14 (Confident)	1	up & down button not working	13/12/16	[Signature]	"



Dr. Elizabeth koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-629 189
 K.K Dist., Tamilnadu

DEPARTMENT OF CONSERVATIVE DENTISTRY AND ENDODONTICS

18


2020 U.G CLINIC

2020

19

Date	Name of the item taken for repairing	Quantity	Nature of complaint	Approximate date of return	Name and signature of technician	Name and signature of HOD/Staff	Date of return	Signature of HOD/Staff
12/2/20	chair no-11	1	light not working	16/2/20	Mr. Saugli Saugli	Srinivas Anjan	19/2/20	[Signature]
5/10/20	chair no-2	1	compressor not working	12/10/20	Mr. Saugli Saugli	Dr. Jeyaraj	15/10/20	[Signature]
13/10/20	chair no-3	1	Three way syringe not working	13/10/20	Mr. Saugli Saugli	Dr. Jeyaraj	16/10/20	[Signature]
5/10/20	chair no-4	1	water line not working	11/10/20	Mr. Saugli Saugli	Dr. Jeyaraj	12/10/20	[Signature]
8/10/20	chair no-7	1	Airbrush and microwave not working	9/10/20	Mr. Saugli Saugli	Dr. Jeyaraj	12/10/20	[Signature]
24/10/20	chair no-12,13	2	chair not working	24/10/20	Mr. Saugli Saugli	Dr. Jeyaraj	25/10/20	[Signature]
5/11/20	chair no-11	1	leakage from aural line	5/11/20	Mr. Saugli Saugli	Dr. Jeyaraj	9/11/20	[Signature]
11/11/20	chair no-9	1	light not working	16/11/20	Mr. Saugli Saugli	Dr. Jeyaraj	18/11/20	[Signature]
2/12/20	chair no-1	1	phantom head-waterline not working	5/12/20	Mr. Saugli Saugli	Dr. Jeyaraj	9/12/20	[Signature]
16/12/20	chair no-3	1	phantom head-leakage	11/12/20	Mr. Saugli Saugli	Dr. Jeyaraj	14/12/20	[Signature]




Dr. Elizabeth Koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-625 181
 K.K. Dist., Tamilnadu

DEPARTMENT OF CONSERVATIVE DENTISTRY AND ENDODONTICS

20 2020 - P.G. CLINIC

Date	Name of the item taken for servicing	Quantity	Nature of complaint	Approximate date of return	Name and signature of Technician	Name and signature of HOD/Staff	Date of return	Signature of HOD/Staff
10/1/20	chair no 5	1	Water leakage	15/2/20	M. Sangli <i>[Signature]</i>	J. H. S. <i>[Signature]</i>	18/1/20	<i>[Signature]</i>
20/1/20	chair no 5	1	Compressor not working	3/3/20	M. Sangli <i>[Signature]</i>	J. H. S. <i>[Signature]</i>	1/3/20	<i>[Signature]</i>

21



Dr. Elizabeth Koshi MDS
Principal
Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam Kulasekharam-629 181
K.K. Dist., Tamilnadu

DEPARTMENT OF CONSERVATIVE DENTISTRY AND ENDODONTICS

2019 - U-G CLINIC

Date	Name of the item taken for servicing	Quantity	Nature of complaint	Approximate date of Return	Name and signature of Technician	Name and signature of HOD/Asst	Date of Return	Signature of HOD/Asst
20/2/19	chair no: 2	1	water line not working	25/2/19	Mr. Saugli <i>Saugli</i>	J. M. Anil <i>J. M. Anil</i>	24/2/19	<i>J. M. Anil</i>
19/2/19	chair no: 3	1	light not working	24/2/19	Mr. Saugli <i>Saugli</i>	Dr. Jey <i>Dr. Jey</i>	24/2/19	<i>Dr. Jey</i>
20/2/19	chair no: 6	1	chair not working	25/2/19	Mr. Saugli <i>Saugli</i>	Dr. Jey <i>Dr. Jey</i>	24/2/19	<i>Dr. Jey</i>
13/2/19	chair no: 10	1	water leakage	18/2/19	Mr. Saugli <i>Saugli</i>	J. M. Anil <i>J. M. Anil</i>	22/2/19	<i>J. M. Anil</i>
20/2/19	chair no: 2	1	airlines not working	25/2/19	Mr. Saugli <i>Saugli</i>	Dr. Jey <i>Dr. Jey</i>	29/2/19	<i>Dr. Jey</i>
12/10/19	chair no: 19	1	airlines not working	21/10/19	Mr. Saugli <i>Saugli</i>	Dr. Jey <i>Dr. Jey</i>	22/10/19	<i>Dr. Jey</i>
19/12/19	chair no: 2	1	airlines leakage and light not working	24/12/19	Mr. Saugli <i>Saugli</i>	J. M. Anil <i>J. M. Anil</i>	4/1/20	<i>J. M. Anil</i>
20/12/19	chair no: 3	1	light not working	25/12/19	Mr. Saugli <i>Saugli</i>	J. M. Anil <i>J. M. Anil</i>	27/12/19	<i>J. M. Anil</i>


Dr. Elizabeth Koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam, Kulasekharam-629 101
 K.K. Dist., Tamilnadu

DEPARTMENT OF CONSERVATIVE DENTISTRY AND ENDODONTICS

2019 - P.G. CLINIC 2019

Date	Name of item taken for repairing	Quantity	Nature of complaint	Approximate date of return	Name and signature of technician	Name and signature of HOD/Staff	Date of Return	Signature of HOD/Staff
15/1/19	chair no: 4	1	light not working	20/1/19	Mr. Sangeetha Sangeetha	S. M. Anand	23/1/19	S. M. Anand
24/1/19	chair no: 4	1	water leakage	29/1/19	Mr. Sangeetha Sangeetha	S. M. Anand	31/1/19	S. M. Anand
15/2/19	chair no: 5	1	compressor complaint	17/2/19	Mr. Sangeetha Sangeetha	S. M. Anand	22/2/19	S. M. Anand
27/2/19	RVA complaint	1	RVA was not working	3/3/19	Mr. Sangeetha Sangeetha	S. M. Anand	21/3/19	S. M. Anand



Dr. Elizabeth Koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-625 181
 K.K. Dist., Tamilnadu

DEPARTMENT OF CONSERVATIVE DENTISTRY AND ENDODONTICS

Date	Name of the Item taken for repairing	Quantity	Nature of complaint	Approximate date of action	Name and signature of Technician	Name and signature of head/dent	Date of action	Signature of
11/1/18	chair no-19	1	3-way syringe (water not coming)	15/1/18	Mr. Sangeetha Sangeetha	J. P. M. Sangeetha	14/1/18	Sangeetha
11/1/18	chair no-18	1	Air motor (water not coming)	15/1/18	Mr. Sangeetha Sangeetha	Dr. Vijay Sangeetha	14/1/18	Sangeetha
11/1/18	chair no-27, 28	1	3-way syringe (air not coming)	16/1/18	Mr. Sangeetha Sangeetha	Dr. Vijay Sangeetha	16/1/18	Sangeetha
12/1/18	chair no-1, 14, 25	3	not working	1/2/18	Mr. Sangeetha Sangeetha	Dr. Vijay Sangeetha	1/2/18	Sangeetha
12/1/18	chair no-13, 21, 24	3	plumbing work	1/2/18	Mr. Sangeetha Sangeetha	Dr. Vijay Sangeetha	1/2/18	Sangeetha
3/2/18	chair no-2	1	handle broken	1/2/18	Mr. Sangeetha Sangeetha	J. P. M. Sangeetha	1/2/18	Sangeetha
12/1/18	chair no-11	1	water not working	2/2/18	Mr. Sangeetha Sangeetha	Dr. Vijay Sangeetha	1/2/18	Sangeetha
12/1/18	chair no-12	1	water leakage	2/2/18	Mr. Sangeetha Sangeetha	Dr. Vijay Sangeetha	2/2/18	Sangeetha
2/2/18	chair no-15, 7, 12	3	light handle broken	20/1/18	Mr. Sangeetha Sangeetha	Dr. Vijay Sangeetha	2/2/18	Sangeetha
24/1/18	chair no-3, 5	2	micromotor, a/c not working	21/1/18	Mr. Sangeetha Sangeetha	Dr. Vijay Sangeetha	21/1/18	Sangeetha
24/1/18	chair no-4	1	Air not working	27/1/18	Mr. Sangeetha Sangeetha	J. P. M. Sangeetha	30/1/18	Sangeetha
10/1/18	chair no-1, 22	2	no water supply in spittoon	2/12/17	Mr. Sangeetha Sangeetha	Dr. Vijay Sangeetha	1/12/17	Sangeetha
11/1/18	chair no-10	1	no water supply in airtax	2/12/17	Mr. Sangeetha Sangeetha	Dr. Vijay Sangeetha	5/12/17	Sangeetha
10/1/18	chair no-19, 22	2	no electricity supply	1/12/17	Mr. Sangeetha Sangeetha	J. P. M. Sangeetha	1/12/17	Sangeetha



Dr. Elizabeth Koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-629 181
 K.K. Dist., Tamilnadu

DEPARTMENT OF CONSERVATIVE DENTISTRY AND ENDODONTICS

2018- P.G CLINIC		2018						
Date	Name of the item taken for repairing	Quantity	Nature of complaint	Approximate date of return	Name and signature of Technician	Name and signature of Hosp/Staff	Date of return	Signature of Hosp/Staff
25/6/18	chair no. 2	1	light not working, micromotor and not working	30/6/18	Mrs. Sangeetha	S. Mas Anand	30/6/18	[Signature]
25/6/18	chair no. 3	1	light not working	30/6/18	Mrs. Sangeetha	S. Mas Anand	30/6/18	[Signature]



Dr. Elizabeth Koshi MDS
 Principal
 Sree Meekambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padaanilam, Kulasekharam-628 181
 K.K. Dist., Tamilnadu

DEPARTMENT OF CONSERVATIVE DENTISTRY AND ENDODONTICS

2017- U.G. CLINIC

6

7

Date	Name of the item taken for servicing	Quantity	Nature of complaint	Approximate date of return	Name and signature of technician	Name and signature of HOD/Staff	Date of return	Signature of HOD/Staff
23/01/18	chair No.10	1	Not working	23/01/18	Mr. Sangli Sangli	J. of ana / Angu	23/01/18	J
23/01/18	chair No.11	1		23/01/18	Mr. Sangli Sangli	Dr. Vijay Dr. Vijay	23/01/18	Dr. Vijay
3/02/18	chair no. 1, 12, 21, 22, 23, 26, 27	7	Not working	1/9/18	Mr. Sangli Sangli	Dr. Vijay Dr. Vijay	01/09/18	Dr. Vijay
31/01/18	chair No. 25, 26, 27, 22, 23, 24, 21, 20, 19	9	Not working	5/01/18	Mr. Sangli Sangli	Dr. Vijay Dr. Vijay	5/01/18	Dr. Vijay
31/01/18	chair no. 12	1	Dist light not working	31/01/18	Mr. Sangli Sangli	Dr. Vijay Dr. Vijay	31/01/18	Dr. Vijay
31/01/18	chair no. 12	1	Foot control	31/01/18	Mr. Sangli Sangli	J. of ana / Angu	31/01/18	J
01/02/18	chair no. 1	1	Not working	5/01/18	Mr. Sangli Sangli	Dr. Vijay Dr. Vijay	05/01/18	Dr. Vijay
01/02/18	chair no. 2	1	Drumming noise	5/01/18	Mr. Sangli Sangli	Dr. Vijay Dr. Vijay	05/01/18	Dr. Vijay
04/02/18	chair no. 18	1	Water leakage	10/01/18	Mr. Sangli Sangli	Dr. Vijay Dr. Vijay	13/01/18	Dr. Vijay
06/02/18	chair no. 3	1	dam movement not working	20/01/18	Mr. Sangli Sangli	Dr. Vijay Dr. Vijay	24/01/18	Dr. Vijay
06/02/18	chair no. 21, 18	2	Water leakage	20/01/18	Mr. Sangli Sangli	S. M. An / J	24/01/18	J
06/02/18	chair no. 19, 18	2	3 way syringe (water not coming)	20/01/18	Mr. Sangli Sangli	Dr. Vijay Dr. Vijay	24/01/18	Dr. Vijay
06/02/18	chair no. 1, 25, 26, 27, 28, 29, 22, 19, 20, 21	10	Not working	20/01/18	Mr. Sangli Sangli	Dr. Vijay Dr. Vijay	24/01/18	Dr. Vijay



Dr. Elizabeth koshi MDS
Principal
Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam Kulasekharam-629 101
K.K. Dist., Tamilnadu

DEPARTMENT OF CONSERVATIVE DENTISTRY AND ENDODONTICS

2017- P.G CLINIC

8

9

Date	Name of the item taken for repairing	Quantity	Nature of complaint	Approximate date of action	Name and signature of Technician	Name and signature of HOD/Staff	Date of action	Signature of HOD/Staff
16/10/13	chair no. 2	1	water leakage, machine not working pressure - leak	20/10/13	Mr. Sangeetha	Dr. Jayaprakash	23/10/13	[Signature]
16/10/13	chair no. 3	1	Micromotor not working up down movement - not working. Air hose stand - broken	20/10/13	Mr. Sangeetha	Dr. Vijay	23/10/13	[Signature]
16/10/13	chair no. 4	1	Micromotor not working Air hose stand - broken	22/10/13	Mr. Sangeetha	Dr. Vijay	25/10/13	[Signature]
16/10/13	chair no. 6	1	Delayed suction with leakage micromotor not working	22/10/13	Mr. Sangeetha	Dr. Jayaprakash	25/10/13	[Signature]
16/10/13	chair no. 7	1	chair movement stuck	21/10/13	Mr. Sangeetha	Dr. Vijay	24/10/13	[Signature]
16/10/13	chair no. 5	1	water leakage	20/10/13	Mr. Sangeetha	Dr. Jayaprakash	23/10/13	[Signature]
16/10/13	X-ray tube	1	Wiring loose	21/10/13	Mr. Sangeetha	Dr. Vijay	25/10/13	[Signature]
16/10/13	chair no. 1	1	Hand not broken	20/10/13	Mr. Sangeetha	Dr. Jayaprakash	23/10/13	[Signature]



Dr. Elizabeth Koshi MDS
Principal
Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam Kulasekharam-625 181
K.K. Dist., Tamilnadu

DEPARTMENT OF CONSERVATIVE DENTISTRY AND ENDODONTICS

2016 - U.G CLINIC

2

3

Date	Name of the item taken for servicing	Quantity	Nature of complaint	Approximate date of return	Name and signature of technician	Name and signature of HOD/Staff	Date of Return	Signature of HOD/Staff
2/1/16	chair no. 23	1	Not working	5/1/16	Mr. Sangli Sangli	S. Manoj Anga Jr	2/1/16	Jr
11/1/16	chair no. 12	1	Dim light not working	14/1/16	Mr. Sangli Sangli	A. Vijay Jr	12/1/16	Jr
19/1/16	chair no: 8, 9	2	water leakage	21/1/16	Mr. Sangli Sangli	A. Vijay Jr	20/1/16	Jr
23/1/16	chair no: 3	1	plumbing work	25/1/16	Mr. Sangli Sangli	A. Vijay Jr	5/2/16	Jr
3/2/16	chair no: 15	1	foot control	10/2/16	Mr. Sangli Sangli	S. Manoj Anga Jr	12/2/16	Jr
15/2/16	chair no: 12, 21, 24, 21, 14	5	not working	18/2/16	Mr. Sangli Sangli	A. Vijay Jr	20/2/16	Jr
25/2/16	chair no: 5	1	Dam movement not working	26/2/16	Mr. Sangli Sangli	S. Manoj Anga Jr	2/3/16	Jr



D. Elizabeth koshi MDS
Principal
Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam Kulasekharam-629 101
K.K Dist., Tamilnadu

DEPARTMENT OF CONSERVATIVE DENTISTRY AND ENDODONTICS

2016 - P.G. CLINIC

4

5

Date	Name of the item taken for receiving	Quantity	Nature of complaint	Approximate date of Rebar	Name and signature of Technician	Name and signature of HOD/Staff	Date of Rebar	Signature of HOD/Staff
2/3/16	Chair no-11	1	micrometer - not working up/down movement	4/3/16	Mr. Sangeetha Sangeetha	J. Manoj Kumar	7/3/16	J. Manoj Kumar
9/3/16	Chair no-6	1	Arrest - not working	11/3/16	Mr. Sangeetha Sangeetha	Dr. Jeyaraj	13/3/16	Dr. Jeyaraj
14/3/16	Chair no-16	1	delayed suction with leakage	14/3/16	Mr. Sangeetha Sangeetha	Dr. Jeyaraj	20/3/16	Dr. Jeyaraj
18/3/16	Chair nos. 2, 3	2	Chair movement stuck	4/4/16	Mr. Sangeetha Sangeetha	Dr. Jeyaraj	9/4/16	Dr. Jeyaraj
26/5/16	Chair no-19	1	Hand rest broken	21/5/16	Mr. Sangeetha Sangeetha	Dr. Jeyaraj	25/5/16	Dr. Jeyaraj
26/5/16	Chair nos. 24	2	not working	20/6/16	Mr. Sangeetha Sangeetha	J. Manoj Kumar	3/9/16	J. Manoj Kumar

(Handwritten signature)




Dr. Elizabeth Koshi MDS
Principal
Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam Kulasekharam-629 161
K.K. Dist., Tamilnadu

DEPARTMENT OF PERIODONTICS

2020

S.No	DATE	NAME OF ITEM TAKEN FOR SERVICE	QTY	COMPLAINT	REASON FOR STOP	STATUS
1)	9/1/2020	USD - P 1102/01 [Wood pecker]	1		Water not coming	17/1/2020 Sanghvi
2)	18/1/2020	Chair 3405/3 [Confident]	1	Power supply not working		1/2/2020 Sanghvi
3)	14/1/2020	Light [V6r clinic]	1	Not working		24/1/2020 R. Jeyaraj




Dr. Elizabeth Koshi MDS
 Principal
 Sree Mookambika Institute of Dental Sciences
 V.P.M. Hospital Complex
 Padanilam Kulasekharam-629 101
 K.K Dist., Tamilnadu

DEPARTMENT OF PERIODONTICS

8				9											
S.No	DATE	Name of item taken for Servicing	Qty	Complaint	Repair on	Sign of	Sign of	S.No	DATE	Name of item taken for Servicing	Qty	Complaint	Repair on	Sign of	Sign of
				Uln	Pln							Uln	P6r		
12	17/1/18	Light [Hed Room]	1	Light	25/1/18	Elb	Alb	11	16/1/19	Light [Uln-clinic]	1	Uln Not working			25/1/18
13	1/9/18	Chair no. 2403/20	1	Hand piece not working	28/9/18	Single		23	25/2/19	Ultrascaler PS-Booster 5110112036	1	Handpiece not working			7/3/19
14	6/10/18	GPS-J 120805 (Woodpecker)	1	Handpiece not working	21/10/18	Single		8)	25/4/19	Chair 3102/20 [Confident]	1	Suction not working			17/5/19
15	11/11/18	UDS-J 102401 (Woodpecker)	1	Water not working properly	24/11/18	Single		4)	10/7/19	Computer	1	CPU Complaint			27/7/19
16	21/11/18	Chair no. 2423/4	1	Suction not working	2/12/18	Single		5)	7/8/19	UDS-P-2110911708 (Woodpecker)	1	Water not coming			15/8/19
17	14/12/18	Fan (Professor room)	1	Fan not working	25/12/18	Single		6)	11/9/19	Chair - 5 [Confident]	1	Airator foot pedal not working			21/9/19
								7)	8/10/19	Fan [Reader room]	1	Not working			28/10/19
								8)	25/10/19	EMS-CE081005 [miniplazon]	1	Handpiece not working			7/11/19
								9)	14/11/19	Fan [Uln-clinic]	1	Not working			29/11/19



Dr. Elizabeth koshi MDS

Principal

Sree Mookambika Institute of Dental Sciences

V.P.M. Hospital Complex

Padanilam Kulasekharam-628 161

K.K. Dist., Tamilnadu

DEPARTMENT OF PERIODONTICS

6				2018				7							
S.No	DATE	Name of item taken for servicing	Qty	Complaint	Repair on	Legal	Sign	S.No	DATE	Name of item taken for servicing	Qty	Complaint	Repair on	Sign of Technician	Sign of Staff
				Un	Pa										
14	06/10/17	Chair no 3228/11 (Confident)	1	Section not working		21/10/17	Sangeetha	1	12/1/18	Chair No - 3402/20	1	X-ray viewer not working	17/1/18	Sangeetha	Staff
15	16/10/17	Fan (Staff room)	1	Not working		21/10/17	Sangeetha	2	16/1/18	Fan (Staff Room)	1	Not working	20/1/18	Ramesh	Staff
16	26/10/17	DDS-J 120505 (WOODPECKER)	1	Hand piece not working		02/11/17	Sangeetha	3	18/1/18	Computer system	1	Keyboard Complaint	22/1/18	Ramesh	Staff
17	11/11/17	DDS-J 120801 (WOODPECKER)	1	Hand piece coming		21/11/17	Sangeetha	4	22/1/18	light (Staff room)	1	Not working	28/1/18	Ramesh	Staff
18	24/11/17	Chair no 3402/10	1	X-ray viewer not working		04/12/17	Sangeetha	5	2/2/18	Chair No - 5 (Confident)	1	Speedies not controlled	13/2/18	Sangeetha	Staff
19	27/11/17	light (Staff Room)	1	Not working		02/11/17	Ramesh	6	6/2/18	DDS-J 122805735 (WOOD PECKER)	1	Hand piece is not working	16/2/18	Sangeetha	Staff
20	08/12/17	Chair no 3405/3	1	Section not working		21/12/17	Sangeetha	7	21/2/18	Fan (Staff Room)	1	Not Working	03/3/18	Ramesh	Staff
21	26/12/17	Fan (Staff room)	1	Not working		31/12/17	Ramesh	8	17/3/18	Switchboard	1	Not working	21/3/18	Ramesh	Staff
								9	4/5/18	ultrascalar unit 809811-019 (Ps booster)	1	Water is not coming	21/5/18	Sangeetha	Staff
								10	22/5/18	Chair No - 11 (Confident)	1	Section not working	29/5/18	Sangeetha	Staff
								11	4/7/18	Computer	1	Have not working	10/7/18	Ramesh	Staff



[Handwritten signature in green ink]

Dr. Elizabeth koshi MDS
Principal
Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam Kulasekharam-629 181
K.K. Dist., Tamilnadu

DEPARTMENT OF PERIODONTICS

Sl No	DATE	Name of item taken for servicing	Qty	Complaint		Signed by or Technician	Date	Sl No	DATE	Name of item taken for servicing	Qty	Complaint	
				UG	P.G							UG	P.G
12	11/11/16	Chair - 1205/1 [Confident]	1	Suction not working		18/11/16	18/11/16	1	15/01/2017	UDS-J 100901 (WOOD PECKER)	1	Handpiece not working	
13	22/11/16	Fan [Staff room]	1	not working		28/11/16	28/11/16	2	20/01/2017	Light (Staff room)	1	Not working	
14	13/11/16	Chair no. 5 [Confident]	1	Airrotor foot pedal not working		27/11/16	27/11/16	3	2/02/17	PS Boxes - Ultrazer Unit 6140112034	1	water not coming properly	
15	15/11/16	Chair 34 05/3 [Confident]	1	X-ray viewer not working		21/11/16	21/11/16	4	6/02/17	Chair no. 3402/1	1	Suction not working	
16	19/12/16	UDS-J 512805735 [Wood pecker]	1	Handpiece is not working		29/12/16	29/12/16	5	28/02/17	Fan [HOD Room]	1	Not working	
								6	18/03/17	Computer System	1	Keyboard complaint	
								7	6/05/17	UDS-P 120601 (WOOD PECKER)	1	Water not coming properly	
								8	16/05/17	UDS-P 120603	1	Speed is not controlled	
								9	3/07/17	Chair no. 13 [Confident]	1	Airrotor foot pedal not working	
								10	18/07/17	Fan [Staff room]	1	Not working	
								11	01/09/17	Light [HOD Room]	1	Not working	
								12	03/09/17	Chair no 3402/8 [Confident]	1	Water supply not working	
								13	24/09/17	Chair no. 7 [Confident]	1	Suction not working	



Handwritten signature in green ink.

Dr. Elizabeth koshi MDS
Principal
Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam, Kulasekharam-626 101
K.K. Dist., Tamilnadu

DEPARTMENT OF PERIODONTICS

2016

S. NO	DATE	Name of item taken for servicing	Qty	COMPLAINT		Repaired on	Sign of Technician	Sign of MOD/LEAD
				UG	PG			
1	20/01/16	UDS-P 518A0113P [WOODPECKER]	1	Water is not coming		02/02/16	Singh	
2	27/01/16	Chair - 3403/1 (Confident)	1	3way Syringe is not working		11/02/16	Singh	Devi
3	04/03/16	Fan [Professor room]	1	NOT WORKING		16/03/16	Devi	Jully
4	22/03/16	UDS-P 512505737 [Woodpecker]	1		Handpiece not working	5/04/16	Singh	Devi
5	13/04/16	Light [MOD ROOM]	1	Not Working		28/04/16	Devi	Devi
6	2/6/16	Chair no. 2	1		Suction not working	13/6/16	Singh	Chand
7	17/6/16	Ultrascaler unit 8098 11 019 [PS Booster]	1	Water is not coming		24/6/16	Singh	Devi
8	06/7/16	Chair - 3402/17 [Confident]	1	Light is not working		21/07/16	Singh	Chand
9	15/7/16	UDS-P 51950348P [Woodpecker]	1		Handpiece not working	26/07/16	Singh	Devi
10	2/9/16	Computer	1		mouse not working	12/09/16	Devi	Devi
11	16/9/16	Scratch board		not working		22/9/16	Devi	Devi



Dr. Elizabeth Koshi MDS
Principal
Sree Mookambika Institute of Dental Sciences
V.P.M. Hospital Complex
Padanilam Kulasekharam-629 101
K.K. Dist., Tamilnadu